

**WAC 246-824-230 Minimum fitting equipment.** Dispensing opticians shall have direct access to the following equipment while fitting contact lenses: Slitlamp or biomicroscope (for evaluation of the fit only), radioscope, diameter gauge, thickness gauge, lensometer, and keratometer.

[Statutory Authority: RCW 18.130.070, 43.17.060 and 43.70.040. WSR 94-06-047, § 246-824-230, filed 3/1/94, effective 4/1/94.]